


Product Name :
Automatic Vacuum Tissue Processor

Product Code :
NLAB-LABORATORYEM28009



Description :

Automatic Vacuum Tissue Processor

Technical Specification :

Steps of processing tissue: 12 steps.

Glass containers: 9 with capacity of 1.2L each.

Wax cups: 3 with capacity of 1L each. Temperature range: 45~85C±3

Process: Each cup working hours between 0-59hr59min. Delay process: 0-59hrs59min.

Vacuum: ? 0.053Mpa.

Vertically reciprocal motion of tissue basket: 3 times/min.

Tissue basket specification: Barrel tissue basket- \varnothing 95x80mm. Layer nacelle- \varnothing 95x80mm, Five layers.

External dimension: (diameter) 660 x (high)550 x 680(Tip top) mm.

Rated voltage: 220V, 50Hz or 110V, 60Hz

Adopting computer to control whole work process, operate brief work stabilization.

Having exchange jar and change place control draw perch, operate agility brief.

If lose electric, alarm sounds to protect tissue in liquid.

Adopting agitation dehydrate mode enhance reagent and paraffin penetration into tissues.

Easy to use with screen display in English.

Wide time range: delay time 59hrs59min. The longest working hours for each container is 59hr59min, Time design least interval is 1 min. Each jar can be set with vacuum separately.

Wax jars adopt constant temperature control system. Temperature precision high: defend pollution, defend canker.

Adopting 9 slices 1.2L medicine aquarium, at any moment observe tissue change at work.

Naugralabequipments

Website: www.naugralabequipments.com, **Email:** sales@naugralabequipments.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India. **Phone:** +91-9896600003