

Product Name : Respironics Lifecare Ventilator	Product Code : NLABMEDICAL-M3390005
Description :	
Respironics Lifecare Ventilator	
Technical Specification :	
Mode Selection: Control, Control + Sigh, Assist/Control, Assist/Control + Sigh, SIMV.	
Tidal Volume: 0.05 to 0.20 $\pm$ 0.02 liters; 0.20 to 3.00 liters $\pm$ 10%.	
Sign Volume: 150% of Set Tidal Volume.	
Sign Frequency: 1 Sign Breath every 100 Tidal.	
Breath Rate: 2 to 35 BPM $\pm$ 0.5; 36 to 40 BPM $\pm$ 2..	
Aspiratory Flow Rate (Peak flow rate during aspiratory time): 10 to 120 LPM.	
Battery Test Switch: Test Internal/External Battery.	
Main Power Switch: On/Off, Recharge.	
DC Circuit Breaker: Push/Pull Actuation.	
AC Circuit Breaker: Push/Pull Actuation.	
Airway Pressure Limit: 5 to 100 cm H2O $\pm$ 5.	
Sensitivity: +18 cm H2O $\pm$ 1 to at least -6 cm H2O.	
Low Pressure Alarm: 2 to 40 cm H2O $\pm$ 2; 41 to 50 cm H2O $\pm$ 5.	
Manual Sigh: Push to Activate One Breath.	

Naugralabequipments

Website: www.naugralabequipments.com, **Email:** sales@naugralabequipments.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India. **Phone:** +91-9896600003