

Product Name : Siemens 900C Ventilator	Product Code : NLABMEDICAL-M3390003
Description :	
Siemens 900C Ventilator	
Technical Specification :	
Gas consumption: Approximately equal to preset aspiratory minute volume or equal to flow meter setting.	
Gas supply via high pressure inlet: From central gas supply, compressor or gas cylinder.	
Electronic gas supply unit, inlet pressure: 2.5-5 bar (250-500 kPa, 35-70 psi).	
Other gas supply inlet pressure: 2.5-7 bar (250-700 kpa, 35-100 psi).	
Gas supply via low pressure inlet: Vial flow meters.	
Power supply: (100), 110, 120, 220, 240, 50 or 60 hz.	
Power consumption: Approximately 40 W.	

Naugralabequipments

Website: www.naugralabequipments.com, **Email:** sales@naugralabequipments.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India. **Phone:** +91-9896600003