

Product Name :
Spectrum Analyzers Testing

Product Code :
SPA-0001



Description :

Spectrum Analyzers Testing

Technical Specification :

Accuracy of Mark Frequency : $\pm 0.1\%$ (frequency sweep width $\pm 100\text{KHz}$)

Frequency Display Resolution : 100KHz (4 1/2 bit LED display)

Frequency Sweep Width : 0.1MHz/div - 100KHz/div with 1, 2 and step system and 0Hz/div (0 frequency sweep)
accuracy : $\pm 10\%$

Intermediate Frequency Bandwidth : 20KHz and 400KHz

Video Filter : (when on) 4KHz

Amplitude Range : $-107 \sim +13\text{dB}$

Reference Level : $-27\text{dB} \sim +13\text{dB}$ (10dB for every step) $\pm 2\text{dB}$

Average Noise Level : -97dB (20KHz intermediate frequency width)

Naugralabequipments

Website: www.naugralabequipments.com, **Email:** sales@naugralabequipments.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India. **Phone:** +91-9896600003